

PPS-20

Series



Precision Piezo Stage
Reference Manual
(Open and Closed Loop Versions)

micronix
PRECISION MOTION SOLUTIONS USA

PPS-20

Piezo Motor

Precision Positioner Stage

Reference Manual

Rev 3.4

MICRONIX USA, LLC
Tel: 949-480-0538
Fax: 949-480-0538
Email: info@micronixusa.com
<http://micronixusa.com>

Table of Contents

1. Introduction.....	3
1.1 Product Description.....	3
1.2 Recommended Controllers.....	3
1.3 Technical Data	3
2. Preparing to Install the PPS-20PM Stage.....	4
2.1 Installation Preparation.....	4
2.2 Package Contents.....	4
3. Installing the PPS-20PM Stage.....	5
3.1 PPS-20PM Installation.....	5
3.1.1 General Mounting	5
3.1.2 X-Y Mounting.....	6
4. Dimensions.....	7
4.1 PPS-20PM Standard Dimensions	7
4.2 Recommended General Mounting Pattern	7
5. Stacking Configurations	8
5.1 Configuration Examples (additional configurations available upon request)	8
5.2 Accessories.....	10
6. Connecting the PPS-20PM Stage.....	11
6.1 Atmospheric Environments.....	11
6.1.1 Open Loop, Atmospheric Wiring Diagram.....	11
6.1.2 Closed Loop (Encoder), Atmospheric Wiring Diagram	12
6.2 Vacuum Environments.....	14
6.2.1 Handling and Preparation	14
6.2.2 Open loop, Vacuum Wiring Diagram.....	15
6.2.3 Closed Loop (Encoder), Vacuum Wiring Diagram	16
7. Supplementary Information	18
7.1 Maintenance	18
7.2 Units and Conventions.....	18
A. Appendix.....	19
A.1 2 Phase Piezo Motor Wiring Diagram	19
A.1.1 Piezo Operating and Electrical Specifications	19
A.2 Feedthroughs.....	19
A.3 Using an Analog Encoder	20
A.3.1 Encoder Overview	20
A.3.2 Encoder Operating and Electrical Specifications	20
A.3.3 Analog Output (Pins 1,2,6, and 7)	20
A.3.4 Index Window	20
A.3.5 Resolution	20
A.4 Using a Digital Encoder.....	21
A.4.1 Encoder Overview	21
A.4.2 Operating and Electrical Specifications	21
A.4.3 Output Signals & Signal Termination for A quad B, Index and limits	21
A.4.4 Resolution	21

A.5	Using an Absolute Encoder.....	22
A.5.1	Encoder Operating and Electrical Specifications	22
A.5.2	Absolute Encoder Pinout.....	22
A.5.3	Absolute Encoder Setup.....	23
A.6	Legacy.....	24
A.6.1	Legacy Digital Encoder with DSub15 Interpolator Module, Atmospheric	24
A.6.2	Legacy Digital Encoder with Dsub15 Interpolator Module, Vacuum.....	25
A.6.1	Legacy Cabling (8 Conductor Wiring)	26
A.6.2	Legacy PPS-20PM with UHV M1000V Analog Encoder Dimensions	27
A.6.3	Legacy Analog M1000V Encoder, Ultra High Vacuum Wiring Diagram	28
A.6.4	Legacy PPS-20PM with UHV MII 6000V Digital Encoder Dimensions	29
A.6.5	Legacy Digital MII6000V Encoder, Ultra High Vacuum Wiring Diagram	30
A.6.6	Legacy UHV Digital Encoder Module	31
A.6.7	Resolution	32

1. Introduction

1.1 Product Description

The PPS-20 is a high-precision, long travel linear piezo stage. Miniature crossed roller bearings assure high stiffness and guiding accuracy for loads up to 20 N (horizontal orientation). It utilizes our patented multi-phase piezo motor resulting in high speed (> 10 mm/s) and high blocking force (> 2 N). The PPS-20 is available in open loop or with an external encoder. Closed loop encoder resolution of 2 nm is achievable. Versions capable of operation in vacuum (10^{-9} mbar), cryogenic (4 Kelvin) and non-magnetic materials are available. The PPS-20 is compatible with the MMC-100, MMC-110 and NanoDrive controllers.

Features:

- Travel range of up to 51 mm
- 2 nm closed loop encoder resolution
- Load capacity up to 2 kg
- Crossed roller bearing
- Low profile, 13 mm height
- Vacuum, cryogenic, and non-magnetic versions available




Figure 1-A. PPS-20PM, Piezo Motor, Open Loop Version (left), Closed Loop Version (right)

1.2 Recommended Controllers

The following controllers are available from MICRONIX USA:

- MMC-100
- MMC-110
- NanoDrive

1.3 Technical Data and Ordering Information

See Datasheet on website.

2. Preparing to Install the PPS-20PM Stage

2.1 Installation Preparation

When mounting the stage, it is important to consider the flatness of the mounting surface, as the stage will conform to the shape of that surface and affect the stage's performance and structural integrity. It is required to have a mounting surface with flatness less than the overall specified flatness of the base.

The stage is calibrated and guaranteed to be within specification at $20^{\circ}\text{C} \pm 5^{\circ}\text{C}$ unless otherwise specified. Be sure to use the stage under the following conditions:

- Mount to a clean surface which is free of debris, burrs or dings with a flatness to be less than the flatness of the base as specified on the product datasheet.
- An indoor atmosphere free of corrosive gases, excessive dust, and condensation.
- Temperature range of 5-40°C.
- Relative humidity between 20-80%.
- Locate away from water, heat, and electrical noise.

2.2 Package Contents

If the product is damaged or there are missing components, contact MICRONIX USA immediately. Do not discard product packaging in case of return shipment.

Package Should Contain:

- PPS-20 Linear Stage
- Reference Manual
- Any other previously agreed upon components such as a controller.

3. Installing the PPS-20PM Stage

Refer to Section 3.1.1 for general mounting, 3.1.2 for XY Mounting. Additional brackets and screws may be required for custom applications, see Section 5 for stacking configuration examples.

3.1 PPS-20PM Installation

3.1.1 General Mounting

Recommended general mounting pattern sample can be found in Section 4.2.

1. Align the stage to the mounting surface using at least two M1.5 dowel pins.
2. Move the carriage to access the mounting holes. Secure the stage to the mounting surface using at least two M2 socket head cap screws at 0.22 Nm recommended torque.

Please note, it is possible to move the carriage of the piezo motor configurations manually without damaging the stage.




Figure 3-A. PPS-20PM Installation

3.1.2 X-Y Mounting

It is possible to mount the 20 mm length base versions directly onto any length bottom axis carriage without the use of an adapter bracket. Contact MICRONIX USA for additional or custom adapter plate information.

1. Install the bottom stage to the mounting surface as shown in Section 3.1.1.
2. Place the top stage and manually move the carriage to access the mounting holes. Secure the stage using two M2 x 4 mm socket head cap screws at 0.22 Nm recommended torque.

Note: Do not use screws longer than specified to avoid damage to the bearings.

Please note, it is possible to move the carriage of the piezo motor configurations manually without damaging the stage.





Figure 3-B. PPS-20PM XY Mounting Installation


4. Dimensions

4.1 PPS-20PM Standard Dimensions



4.2 Recommended General Mounting Pattern


It is recommended to use a pin-slot hole pattern for dowel pin alignment.




5. Stacking Configurations

5.1 Configuration Examples (additional configurations available upon request)


- Additional configurations available upon request
- Positioning according to:




No Adapters



XY 12mm x 12mm




XY 12mm x 12mm Back-to-Back




XY 12mm x 18mm

Using Z Bracket PN: 430509




XYZ 12mm x 12mm x 12mm




XYZ 51mm x 12mm x 26mm

Using Z Bracket PN: 430508 and XY Bracket PN: 431331




Using Z Bracket PN: 430507 and XY Bracket PN: 431331




5.2 Accessories

430507
Z-Bracket Adapter




Used to adapt the PPS-20 series to a perpendicular assembly for XZ mounting configurations.

430508
Z-Bracket Adapter




Used to adapt the PPS-20 series to a perpendicular assembly for XZ mounting configurations.

430509
Z-Bracket Adapter



Used to adapt the PPS-20 series to a perpendicular assembly for XZ mounting configurations.

431331
XY Bracket



Used to adapt the PPS-20 series to an assembly for a XY mounting configurations.

Note: For longer or custom brackets contact MICRONIX USA

6. Connecting the PPS-20PM Stage

6.1 Atmospheric Environments

For controller information, refer to the appropriate MMC controller manual.

6.1.1 Open Loop, Atmospheric Wiring Diagram

Connecting the PPS-20PM in an open loop configuration only requires that the D-sub 9 Pin Motor Cable be connected to a compatible controller. No other cables or components are required.

Cable Descriptions:

- A. Motor Cable (Male Dsub9 Pin, 1.5m PVC Black Cable)





Figure 6-A. PPS-20PM, Piezo Motor, Open Loop, Atmospheric Wiring Diagram

6.1.1.1 Piezo Motor Atmospheric Open Loop Pinout

Pinout for PPS-20-1X000			Cable A Dsub9M
Description	Color	L1	
Phase 1	Red	1	
Phase 2	White (Green TP)	2	
Ground	Black/Green	5	
Shield	-	Casing	
Limit Switch -	White (Violet TP)	6	
Limit Switch +	Violet	7	
+5V	White (Grey)	8	
GND	Grey	9	

Motor
Limit Switch (Optional)



Dsub9M - Front View
9 Pin Male Connector

6.1.2 Closed Loop (Encoder), Atmospheric Wiring Diagram

Using the PPS-20PM stage with an encoder requires a closed loop compatible controller that recognizes the proper type of encoder feedback. Connect the stage as shown below.

Cable Descriptions:

- A. Motor Cable (Male Dsub9 Pin, 1.5m PVC Black Cable)
- B. Encoder Cable (Female Dsub9 Pin, 1.5m PVC Black Cable)






Figure 6-B. PPS-20PM, Piezo Motor, Closed Loop, Atmospheric Wiring Diagram

6.1.2.1 Piezo Motor Atmospheric Analog Encoder Pinout

Pinout for PPS-20-1X200			Cable A Dsub9M			Pinout for PPS-20-1X200			Cable B Dsub9F		
Description	Color	L1	Description	Color	L2	Description	Color	L1	Description	Color	L2
Phase 1	Red	1	Cos+	Blue	1	Index+	Violet	3	Index-	White (Violet TP)	8
Phase 2	White (Green TP)	2	Sin+	Brown	2	GND	Grey	4	Shield	-	Casing
Ground	Black/Green	5	+5V	White (Grey TP)	5	Cos-	White (Blue TP)	6			
Shield	-	Casing	Sin-	White (Brown TP)	7						



Dsub9M - Front View
9 Pin Male Connector



Dsub9F - Front View
9 Pin Female Connector

6.1.2.2 Piezo Motor Atmospheric Digital Encoder Pinout


MOTOR

Pinout for PPS-20-1X300		
Description	Color	L1
Phase 1	Red	1
Phase 2	White (Green TP)	2
Ground	Black/Green	5
Shield	-	Casing


Cable A
Dsub9M

Pinout for PPS-20-1X300		
Description	Color	L2
A+	Blue	1
B+	Brown	2
Index+	Violet	3
GND	Grey	4
+5V	White (Grey TP)	5
A-	White (Blue TP)	6
B-	White (Brown TP)	7
Index-	White (Violet TP)	8
Shield	-	Casing

Encoder



Dsub9M - Front View
9 Pin Male Connector



Dsub9F - Front View
9 Pin Female Connector

6.2 Vacuum Environments

6.2.1 Handling and Preparation

When handling the stage for vacuum environments, take the necessary precautions, such as wearing clean latex gloves, clean room, clothing, etc. Avoid any contaminants.

Maximum bake-out temperature is 100°C. MICRONIX USA optionally supplies the stage with vacuum compatible connectors, see chart below.

Connector Description	Connector Material	Contacts	Backshell
High Vacuum Glass-filled Dyathilate D-Subminiature	DAP	T2 Female Crimps, Gold Pins (Accuglass P/N: 111652)	Nickle-plated Zinc Backshell Strain Relief
Ultra High Vacuum D-Subminiature	PEEK	T1 Female Crimps, Gold Pins (Accuglass P/N: 100180)	PEEK UHV Strain Relief

Environment	Open Loop	Closed Loop
High Vacuum (10^{-6} mbar)	9 Pin Female DAP	15 Pin Female DAP
Ultra-High Vacuum (10^{-9} mbar)	9 Pin Female PEEK	15/25 Pin Female PEEK

Connecting an open loop PPS-20PM in a vacuum chamber requires the use of a feed-through connector at the vacuum chamber wall.

The vacuum compatible PPS-20PM will be supplied with wiring for a straight through feed-through, not a cross over gender changer. MICRONIX USA supplies test connectors that simulate the vacuum feed-through to allow for functionality testing prior to installation in a vacuum chamber, see Appendix A.2 for feedthrough pins.

6.2.2 Open loop, Vacuum Wiring Diagram

Cable Descriptions:

- A. Vacuum Motor Cable (Female Dsub9 Pin Peek or DAP, 1.5m Silver Braided Cable)
- B. Atmospheric Motor Cable (Female Dsub9 Pin to Male Dsub9 Pin, 1.5m PVC Black Cable)





Figure 6-C. PPS-20PM, Piezo Motor, Open Loop, Vacuum Wiring Diagram


6.2.2.1 Piezo Motor Vacuum Open Loop Pinout

Pinout for PPS-20-1X006/1X009		Cable A Dsub9F			Feedthrough Dsub9M		Cable B Dsub9F Dsub9M	
Description	Color	L1	L2	L3	Color	L4	L5	
Phase 1	Red	5	5	1	Red	1	1	
Phase 2	White (Green TP)	4	4	2	White (Green TP)	2	2	
Ground	Black/Green	1	1	5	Black/Green	5	5	
Limit Switch -	White (Violet TP)	9	9	6	White (Violet TP)	6	6	
Limit Switch +	Violet	8	8	7	Violet	7	7	
+5V	White (Grey)	7	7	8	White (Grey)	8	8	
GND	Grey	6	6	9	Grey	9	9	

MOTOR
Limit
Switch
(Optional)



Dsub9F - Front View
9 Pin Female Connector



Dsub9M - Front View
9 Pin Male Connector

6.2.3 Closed Loop (Encoder), Vacuum Wiring Diagram

Cable Descriptions:

- A. Vacuum Motor Cable (Female Dsub15 Peek or DAP, 1.5m Silver Braided Cable)
- B. Vacuum Encoder Cable (Female Dsub15 Peek or DAP, 1.5m Silver Braided Cable)
- C. Atmospheric Motor Breakout Cable (Female Dsub 15 Pin to Male Dsub 9 Pin, 1.5m PVC Black Cable)
- D. Atmospheric Encoder Breakout Cable (Female Dsub 15 Pin to Female Dsub 9 Pin, 1.5m PVC Black Cable)






Figure 6-D. PPS-20PM, Piezo Motor, Closed Loop, Vacuum Wiring Diagram

6.2.3.1 Piezo Motor Vacuum Analog Encoder Pinout


Pinout for PPS-20-1X206/1X209		Cable A&B Dsub15F			Feedthrough Dsub15M			Dsub15F			Cable C Dsub9M	Cable D Dsub9F
Description	Color	L1	L2	L3	Color	L4	L5	L6				
Phase 1	Red	1	1	8	Red	8	1	-				
Phase 2	White (Green TP)	2	2	7	White (Green TP)	7	2	-				
Ground	Black/Green	9	9	15	Black/Green	15	5	-				
Shield	-	10	10	14	-	14	Casing	-				
Cos+	Blue	7	7	2	Blue	2	-	1				
Sin+	Brown	4	4	5	Brown	5	-	2				
Index+	Violet	14	14	10	Violet	10	-	3				
GND	Grey	8	8	1	Grey	1	-	4				
+5V	White (Grey TP)	6	6	3	White (Grey TP)	3	-	5				
Cos-	White (Blue TP)	5	5	4	White (Blue TP)	4	-	6				
Sin-	White (Brown TP)	12	12	12	White (Brown TP)	12	-	7				
Index-	White (Violet TP)	13	13	11	White (Violet TP)	11	-	8				
Shield	-	15	15	9	-	9	-	Casing				



Dsub15F - Front View
15 Pin Female Connector




Dsub9M - Front View
9 Pin Male Connector




Dsub9F - Front View
9 Pin Female Connector

6.2.3.2 Piezo Motor Vacuum Digital Encoder Pinout


Pinout for PPS-20-1X306/1X309		Cable A&B Dsub15F			Feedthrough Dsub15M			Dsub15F			Cable C Dsub9M	Cable D Dsub9F
MOTOR <i>A & C</i>	Encoder <i>B & D</i>	Description	Color	L1	L2	L3	Color	L4	L5	L6		
		Phase 1	Red	1	1	8	Red	8	1	-		
		Phase 2	White (Green TP)	2	2	7	White (Green TP)	7	2	-		
		Ground	Black/Green	9	9	15	Black/Green	15	5	-		
		Shield	-	10	10	14	-	14	Casing	-		
		A+	Blue	7	7	2	Blue	2	-	1		
		B+	Brown	4	4	5	Brown	5	-	2		
		Index+	Violet	14	14	10	Violet	10	-	3		
		GND	Grey	8	8	1	Grey	1	-	4		
		+5V	White (Grey TP)	6	6	3	White (Grey TP)	3	-	5		
		A-	White (Blue TP)	5	5	4	White (Blue TP)	4	-	6		
		B-	White (Brown TP)	12	12	12	White (Brown TP)	12	-	7		
		Index-	White (Violet TP)	13	13	10	White (Violet TP)	10	-	8		
		Shield	-	15	15	9	-	9	-	Casing		



Dsub15F - Front View
15 Pin Female Connector



Dsub9M - Front View
9 Pin Male Connector



Dsub9F - Front View
9 Pin Female Connector

7. Supplementary Information

7.1 Maintenance

- The PPS-20 linear stage series utilizes a maintenance free design. Do not modify the stage or perform any maintenance unless specifically instructed to do so by MICRONIX USA personnel. If the stage is not performing up to the original specifications, please contact MICRONIX USA.
- The PPS-20 series of linear stages are precision mechanical devices and should be handled with care. Do not drop or mishandle the stage.
- Do not touch the scale, as this will contaminate and jeopardize the performance of the stage.
- Follow the *Installation Preparation* requirements and use proper cable management to ensure a clean and safe operating environment.


7.2 Units and Conventions

All measurements in this document are in the metric system of units.

Metric Unit	English Unit
1 millimeter	0.0394 inches
1 micron	0.0000394 inches
1 Newton	0.2248 lbs.
1 Newton-meter	8.85 in-lbs.

A. Appendix

A.1 2 Phase Piezo Motor Wiring Diagram




A.1.1 Piezo Operating and Electrical Specifications


Voltage	60V maximum
Capacitance	150nf ±15%

A.2 Feedthroughs

Straight Through 9-Pin Feed-through



Straight Through 15-Pin Feed-through



A.3 Using an Analog Encoder


A.3.1 Encoder Overview

A PPS-20 with analog encoder will need to be paired with an appropriate controller. The PPS-20 with an analog encoder will be supplied with a 9-pin connector.


A.3.2 Encoder Operating and Electrical Specifications

Power Supply	5VDC $\pm 10\%$ (< 50mA for sensor)
Storage Temperature	-20 °C to +70 °C
Operating Temperature	0 °C to +70 °C
Humidity	95% relative humidity (non-condensing)

A.3.3 Analog Output (Pins 1,2,6, and 7)



A.3.4 Index Window



A.3.5 Resolution

All closed loop stages are supplied with 20µm scales. The interpolation is done in the MMC-100 to a higher resolution as specified in the order. With an analog encoder the MMC-100 has an achievable resolution of 10nm.

A.4 Using a Digital Encoder


A.4.1 Encoder Overview

A PPS-20 with Digital Encoder will need to be paired with an appropriate controller. The MMC-100 has a Digital option. The PPS-20 with a digital encoder will be supplied with a 9-pin connector.

A.4.2 Operating and Electrical Specifications

Power Supply	5VDC ±10% @ < 35mA (No outputs terminated) @ < 85mA (A, B, I, and both limits terminated); 50mA at the sensor
--------------	--

A.4.3 Output Signals & Signal Termination for A quad B, Index and limits



*Output signals are differential. Inverse signals are not shown for clarity.

Note: At some interpolations values the index pulse may be aligned with other states of A or B than the ones shown.


Above are with reference to the sensor's optical centerline

A.4.4 Resolution

All closed loop stages are supplied with 20µm scales. The digital encoder module interpolates to a higher resolution as specified in the order. With a digital encoder an MMC controller has an achievable resolution of 2nm.

A.5 Using an Absolute Encoder

The Micronix absolute encoder operates using standard BiSS C-mode (continuous) interface, transmitting 32-bits of position data on each request. The controller will clock position acquisitions via the MA signal. The SLO signal will transmit position data from the encoder.




A.5.1 Encoder Operating and Electrical Specifications

Power Supply	5VDC ±10% (< 30mA for sensor)
--------------	-------------------------------

A.5.2 Absolute Encoder Pinout

Pin Dsub9F	Description
1	SLO+ / DATA+
2	MA+ / CLK+
3	SLI+
4	Ground
5	+5V
6	SLO- / DATA-
7	MA- / CLK-
8	SLI-
9	Not In Use



Dsub9F - Front View
9 Pin Female Connector

A.5.3 Absolute Encoder Setup

Absolute Encoder BiSS-C	
Absolute Resolution	32 Bits
Encoder BiSS Frequency	5 MHz
Encoder BiSS CRC Polynomial	0
Position Integer Type	Unsigned
Number of Status Bits	0
Error Bits Mask	0
CRC Error Suppression	None
Data Alignment	Left Justified

A.6 Legacy

A.6.1 Legacy Digital Encoder with DSub15 Interpolator Module, Atmospheric

Cable Descriptions:

- A. Motor Cable (Male Dsub9 Pin, 1.5m PVC Black Cable)
- B. Encoder Cable (Male Dsub15 Pin Module, 1m PVC Black Cable)
- C. Encoder Module Adapter Cable (Female Dsub 15 Pin to Female Dsub 9 Pin, 0.5m PVC Black Cable)





Figure A-A. PPS-20PM, Piezo Motor, Digital Encoder, Atmospheric Wiring Diagram

MOTOR


Pinout for PPS-20-1X300

Description	Color	L1
Phase 1	Red	1
Phase 2	White (Green TP)	2
Ground	Black/Green	5
Shield	-	Casing

Cable A
Dsub9M



Dsub15F - Front View
15 Pin Female Connector




Dsub9M - Front View
9 Pin Male Connector

Pinout for PPS-20-1X300

Description	Color	Interpolator		
		Dsub15M	Dsub15F	Dsub9F
A+	Brown	14	14	1
B+	Blue	13	13	2
Index+	Violet	12	12	3
GND	Grey	2	2	4
+5V	White (Grey TP)	7	7	5
A-	White (Brown TP)	6	6	6
B-	White (Blue TP)	5	5	7
Index-	White (Violet TP)	4	4	8
Shield	-	Casing	Casing	Casing

Cable C




Dsub9F - Front View
9 Pin Female Connector

A.6.2 Legacy Digital Encoder with Dsub15 Interpolator Module, Vacuum


Cable Descriptions:

- Vacuum Motor Cable (Female Dsub 15 Pin DAP, 1.5m Silver Braided Cable)
- Vacuum Encoder Cable (Female Dsub 15 Pin DAP, 1.5m Silver Braided Cable)
- Atmospheric Motor Breakout Cable (Female Dsub 15 Pin to Male Dsub 9 Pin, 1.5m PVC Black Cable)
- Atmospheric Encoder Module Breakout Cable (Female Dsub 15 Pin to Interpolator Module, 1m PVC Black Cable)
- Encoder Module Adapter Cable (Female Dsub 15 Pin to Female Dsub 9 Pin, 0.5m PVC Black Cable)




MOTOR A & D	Pinout for PPS-20-1X306			Feedthrough			Cable C Dsub15F Dsub9M	
	Description	Color	L1	L2	L3	Color	L4	L8
Phase 1	Red		1	1	8	Red	8	1
Phase 2	White (Green TP)		2	2	7	White (Green TP)	7	2
Ground	Black/Green		9	9	15	Black/Green	15	5
Shield	-		10	10	14	-	14	Casing <i>To Interpolator L5 (Cable D)</i>
GND	Grey		8	8	1	Grey	1	
Cos+	Blue		7	7	2	Blue	2	
+5V	White (Grey TP)		6	6	3	White (Grey TP)	3	
Cos-	White (Blue TP)		5	5	4	White (Blue TP)	4	
Sin+	Brown		4	4	5	Brown	5	
Sin-	White (Brown TP)		12	12	12	White (Brown TP)	12	
Index-	White (Violet TP)		13	13	11	White (Violet TP)	11	
Index+	Violet		14	14	10	Violet	10	
Shield	-		15	15	9	-	9	


Encoder	Pinout for PPS-20-1X306			Interpolator			Cable E	
	Description	Color	L5	L6	L7	Dsub15M	Dsub15F	Dsub9F
A+	Brown		14	14	1			
B+	Blue		13	13	2			
Index+	Violet		12	12	3			
GND	Grey		2	2	4			
+5V	White (Grey TP)		7	7	5			
A-	White (Brown TP)		6	6	6			
B-	White (Blue TP)		5	5	7			
Index-	White (Violet TP)		4	4	8			
Shield	-		Casing	Casing	Casing			



Dsub15F - Front View
15 Pin Female Connector



Dsub9M - Front View
9 Pin Male Connector




Dsub9F - Front View
9 Pin Female Connector

A.6.1 Legacy Cabling (8 Conductor Wiring)

Stages delivered before 2022 used the 8 Conductor wiring scheme.


Stages delivered on 2022 and afterward use the 12 Conductor wiring scheme.

Customers can find the serial number of their stages which can usually be found on the Dsub 9 Male connector with the format: 1234-56789 where the first four digit is the year.



Description	8 Conductor Wiring	12 Conductor Wiring
Motor Phase 1	Red	Red
Motor Phase 2	Yellow	White (Green TP)
Motor Ground	Black & Green	Black & Green
A+ / COS+	Brown	Blue
A- / COS-	Orange	White (Blue TP)
B+ / SIN+	Yellow	Brown
B- / SIN-	Green	White (Brown TP)
Index +	Violet	Violet
Index -	Blue	White (Violet TP)
+5V	Red	White (Grey TP)
GND	Black	Grey

A.6.2 Legacy PPS-20PM with UHV M1000V Analog Encoder Dimensions



A.6.3 Legacy Analog M1000V Encoder, Ultra High Vacuum Wiring Diagram

Cable Descriptions:

- A. Vacuum Motor Cable (Female Dsub15 Pin Peek, 1.5m Silver Braided Cable)
- B. Vacuum Encoder Cable (Female Dsub15 Pin Peek, 1.5m Silver Braided Cable)
- C. Atmospheric Motor Breakout Cable (Female Dsub15 Pin to Male Dsub 9 Pin, 1.5m PVC Black Cable)
- D. Atmospheric Encoder Breakout Cable (Female Dsub15 Pin to Female Dsub 9 Pin, 1.5m PVC Black Cable)






Figure A-B. PPS-20PM, Piezo Motor, Analog Encoder, UHV Wiring Diagram


Pinout for PPS-20-1X209	Cable A&B Dsub15F			Feedthrough Dsub15M			Dsub15F	Cable D Dsub9F	Cable C Dsub9M
	Description:	Color	L1	L2	L3	Color	L4	L5	L6
Phase 1	Red		1	1	8	Red	8	-	1
Phase 2	White (Green TP)		2	2	7	White (Green TP)	7	-	2
GND	Black/Green		9	9	15	Black/Green	15	-	5
Shield	-		10	10	14	-	14	-	Casing
GND	Black		8	8	1	Grey	1	4	-
Cos+	Brown		7	7	2	Brown	2	1	-
+5V	Red		6	6	3	White (Grey TP)	3	5	-
Cos-	Orange		5	5	4	White (Brown TP)	4	6	-
Sin+	Yellow		4	4	5	Blue	5	2	-
Sin-	Green		12	12	12	White (Blue TP)	12	7	-
Index-	Blue		13	13	11	White (Violet TP)	11	8	-
Index+	Violet		14	14	10	Violet	10	3	-
Shield	-		15	15	9	-	9	Casing	-



Female PEEK Dsub15 Connector - Rear View




Female Dsub9 Connector - Rear View



Male Dsub9 Connector - Rear View

A.6.4 Legacy PPS-20PM with UHV MII 6000V Digital Encoder Dimensions



* 12 mm travel version shown


* all dimensions are in millimeters

* grey parts for closed loop version only

A.6.5 Legacy Digital MI6000V Encoder, Ultra High Vacuum Wiring Diagram

Cable Descriptions:






- Vacuum Motor Cable (Female Dsub 25 Pin DAP, 1.5m Silver Braided Cable)
- Vacuum Encoder Cable (Female Dsub 25 Pin DAP, 1.5m Silver Braided Cable)
- Atmospheric Motor Breakout Cable (Female Dsub 25 Pin to Male Dsub 9 Pin, 1.5m PVC Black Cable)
- Atmospheric Encoder Module Breakout Cable (Female Dsub 25 Pin to Interpolator Module, 1m PVC Black Cable)
- Encoder Module Adapter Cable (Female Dsub 15 Pin to Female Dsub 9 Pin, 0.5m PVC Black Cable)



Pinout for PPS-20-1X309		Cable A&B Dsub25F			Feedthrough Dsub25M		Dsub25F		Cable C Dsub9M	
Description:	Color	L1	L2	L3	Color	L4	L8			
Phase 1	Red	1	1	13	Red	13	1	To Interpolator-L5 (Cable D)		
Phase 2	White (Green TP)	2	2	12	White (Green TP)	12	2			
Ground	Black/Green	14	14	25	Black/Green	25	5			
Shield	-	15	15	24	-	24	Casing			
+5V	Red	4	4	10	Red	10				
GND	Black	17	17	22	Black	22				
DCLK-	Grey	5	5	9	Grey	9				
DCLK+	White (Grey TP)	18	18	21	White (Grey TP)	21				
MISO-	Violet	6	6	8	Violet	8				
MISO+	White (Violet)	19	19	20	White (Violet TP)	20				
MOSI-	Blue	7	7	7	Blue	7				
MOSI+	White (Blue TP)	20	20	19	White (Blue TP)	19				
nSS-	Green	8	8	6	Green	6				
nSS+	White (Green TP)	21	21	18	White (Green TP)	18				
CLK-	Brown	9	9	5	Brown	5				
CLK+	White (Brown TP)	22	22	17	White (Brown TP)	17				
Shield	-	16	16	23	-	23				

Straight Through 25-Pin Feed-through

Pinout for PPS-20-1X309		Interpolator			Cable E		
Description:	Color	Dsub15M	Dsub15F	Dsub9F			
A+	Brown	14	14	1			
B+	Blue	13	13	2			
Index+	Violet	12	12	3			
GND	Grey	2	2	4			
+5V	White (Grey TP)	7	7	5			
A-	White (Brown TP)	6	6	6			
B-	White (Blue TP)	5	5	7			
Index-	White (Violet TP)	4	4	8			
Shield	-	Casing	Casing	Casing			

Dsub25F - Front View
25 Pin Female ConnectorDsub15F - Front View
15 Pin Female ConnectorDsub9M - Front View
9 Pin Male ConnectorDsub9F - Front View
9 Pin Female Connector

A.6.6 Legacy UHV Digital Encoder Module

When using the digital external encoder configuration, the Encoder Module should display two green LED's indicating a power source and proper encoder alignment. A Red or Yellow Signal Level LED indicates misalignment of the Encoder Head, if this occurs contact MICRONIX USA. Do not adjust the Encoder Head or scale. For more information refer to MicroE Systems Mercury Encoders.


A.6.6.1 Legacy UHV Encoder Module Pin-out

Pin	Description	Pin	Description
1	*Right Limit+	9	Ground
2	Ground	10	*Left Limit+
3	*Right Limit-	11	*Left Limit-
4	Index-	12	Index+
5	B-	13	B+
6	A-	14	A+
7	+5V	15	(not used)
8	+5V		

*-Limits must be specified at the time of order and calibrated at the factory.

Note: Tri-state alarm: A and B are tri-stated if the encoder signal becomes too low for reliable operation.


Normal Operation



A.6.6.2 Operating and Electrical Specifications

Power Supply	5VDC $\pm 5\%$ @ 140mA (No outputs terminated) @ 180mA (A, B, I, and both limits terminated); 50mA at the sensor
--------------	---

A.6.6.3 Output Signals & Signal Termination for A quad B, Index and limits



*Output signals are differential. Inverse signals are not shown for clarity.

Note: At some interpolations values the index pulse may be aligned with other states of A or B than the ones shown.

Above are with reference to the sensor's optical centerline

A.6.7 Resolution

All closed loop stages are supplied with 20 μ m scales. The digital encoder module interpolates to a higher resolution as specified in the order. With a digital encoder an MMC controller has an achievable resolution of 2nm.